

ELEKTROTECHNICKÝ ZKUŠEBNÍ ÚSTAV



ELECTROTECHNICAL TESTING INSTITUTE - CZECH REPUBLIC
ELEKTROTECHNISCHE PRÜFANSTALT - TSCHHECHISCHE REPUBLIK
INSTITUT ELECTROTECHNIQUE D'ESSAIS - RÉPUBLIQUE TCHÈQUE
ЭЛЕКТРОТЕХНИЧЕСКИЙ ИСПЫТАТЕЛЬНЫЙ ИНСТИТУТ - ЧЕШСКАЯ РЕСПУБЛИКА

Pod lisem 129/2, 171 02 Praha 8 - Troja

CERTIFICATE

No.: 1210690

Product: Magnetotherapeutic/Stimulation device BioTorus, Dia Torus

Type: LT-100 + FW8001M/12 with applicators A1C, A4C, B1C, ASE and APE

Rating: 100 – 240 V~, 50 - 60 Hz, 400 - 200 mA

Ordering firm: 2EL, spol. s r.o.
Na Staré Cidlině 663, 504 01 Nový Bydžov, Czech Republic

Manufacturer: 2EL, spol. s r.o.
Na Staré Cidlině 663, 504 01 Nový Bydžov, Czech Republic

Factories: 2EL, spol. s r.o.
Na Staré Cidlině 663, 504 01 Nový Bydžov, Czech Republic

Trade mark:

The test results are stated in the test-report No.: 212835-01/01 of: 10.12.2021, 212835-01/02 of: 14.12.2021

A sample of the product was found to be in conformity with:
EN 60601-1:2006+A1:2013, EN 60601-1-2:2015+A1:2021,
EN 60601-1-6:2010+A1:2015, EN 60601-1-11ed.2:2015

Other data:

The validity of the certificate is limited to: 16.12.2024

17.12.2021

Prague

Mgr. Miroslav Sedláček
Head of Certification Body



* C E R / 1 2 1 0 6 9 0 *

212835-01