

ELEKTROTECHNICKÝ ZKUŠEBNÍ ÚSTAV



ELECTROTECHNICAL TESTING INSTITUTE - CZECH REPUBLIC
ELEKTROTECHNISCHE PRÜFANSTALT - TSCHJECHISCHE REPUBLIK
INSTITUT ELECTROTECHNIQUE D'ESSAIS - RÉPUBLIQUE TCHÉQUE
ЭЛЕКТРОТЕХНИЧЕСКИЙ ИСПЫТАТЕЛЬНЫЙ ИНСТИТУТ - ЧЕШСКАЯ РЕСПУБЛИКА

Pod lisem 129/2, 171 02 Praha 8 - Troja

CERTIFICATE

No.: 1210551

Product: Magnetotherapeutic/Stimulating Device UltiCare

Type: LT-99 + adapter FW8001M/12
with applicators A1C, A4C, B1C, ASE and APE

Rating: 100-240 V, 50-60 Hz, 400-200 mA

Ordering firm: 2EL, spol. s r.o.
Na Staré Cidlině 663, 504 01 Nový Bydžov, Czech Republic

Manufacturer: 2EL, spol. s r.o.
Na Staré Cidlině 663, 504 01 Nový Bydžov, Czech Republic

Factories: 2EL, spol. s r.o.
Na Staré Cidlině 663, 504 01 Nový Bydžov, Czech Republic

Trade mark:

The test results are stated in the test-report No.: 212128-01/01 of: 17.09.2021, 212128-01/02 of: 11.10.2021

A sample of the product was found to be in conformity with:
EN 60601-1:2006+A1:2013, EN 60601-1-2:2015+A1:2021, EN 60601-1-6:2010+A1:2015, EN 60601-1-11ed.2:2015

Other data:

The validity of the certificate is limited to: 12.10.2024

13.10.2021

Prague

Mgr. Miroslav Sedláček
Head of Certification Body



212128-01